


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 2188, Rack A, Plate 1, Well E12, 486272
[DC6CC5C3287F933F] Jul 01, 2026 12:55:48 Pressure = 79 bar (79 to 81)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.518	F 0.62	2.0%	19549	
2	0.695	F 0.83	11.8%	117680	Acetylated F peak
3	0.827	F 0.99	73.0%	729725	Consistent with F
4	1.153	A 0.99	13.3%	132464	A peak

Total Area: 999418

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range