


LABORATORIO PREGEN
 Carrera 15a No 106-42
 BOGOTA

Batch 1967, Rack A, Plate 1, Well C12, 170906
 [9C23E07F9927F19D] Nov 05, 2025 13:27:05 Pressure = 64 bar (64 to 66)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.393	F 0.45	1.5%	12320	
2	0.498	F 0.57	1.0%	8439	
3	0.730	F 0.84	17.3%	142757	
4	0.858	F 0.99	51.3%	423203	Consistent with F
5	1.015	F 1.17	2.3%	18949	
6	1.240	A 1.00	26.6%	219976	A peak - REVIEW
Total Area:				825644	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range