


LABORATORIO PREGEN  
Carrera 15a No 106-42  
BOGOTA

Batch 1870, Rack A, Plate 1, Well B08, 458032  
[997425B0A58E511F] Jul 23, 2025 09:56:48 Pressure = 61 bar (60 to 62)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.405	F 0.49	2.5%	40857	
2	0.487	F 0.58	3.5%	57274	
3	0.705	F 0.84	12.6%	205298	Acetylated F peak
4	0.835	F 1.00	60.4%	981049	Consistent with F
5	1.222	A 1.01	20.9%	340246	A peak - REVIEW

Total Area: 1624724

- Codes:
- 1) Wide A peak
  - 2) Area of A peak < 80%
  - 3) Peak area greater than expected
  - 4) Peak after A2
  - 5) Alc > 10%
  - 6) HbF or variant present
  - 7) Total sample area too small/big
  - 8) A2 is not within normal range