


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 2188, Rack A, Plate 1, Well E10, 486270
[883880967D0044CA] Jul 01, 2026 12:49:44 Pressure = 79 bar (79 to 82)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.513	F 0.61	1.5%	18000	
2	0.683	F 0.82	11.1%	137367	Acetylated F peak
3	0.822	F 0.98	53.1%	655878	Consistent with F
4	1.163	A 1.01	34.3%	424187	A peak - REVIEW
Total Area:				1235432	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range