


LABORATORIO PREGEN  
Carrera 15a No 106-42  
BOGOTA

Batch 2073, Rack A, Plate 1, Well H07, 489559  
[8867754ACBAE0C48] Feb 13, 2026 15:39:45 Pressure = 50 bar (50 to 52)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.405	F 0.45	1.2%	15138	
2	0.517	F 0.57	0.4%	5720	
3	0.737	F 0.82	13.6%	173236	Acetylated F peak
4	0.880	F 0.98	66.7%	849506	Consistent with F
5	1.270	A 1.01	18.1%	230469	A peak
Total Area:				1274069	

- Codes:
- 1) Wide A peak
  - 2) Area of A peak < 80%
  - 3) Peak area greater than expected
  - 4) Peak after A2
  - 5) Alc > 10%
  - 6) HbF or variant present
  - 7) Total sample area too small/big
  - 8) A2 is not within normal range