


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 2047, Rack A, Plate 1, Well B04, 490000-2MX
[C837C83D66688EE8] Jan 16, 2026 11:21:33 Pressure = 85 bar (83 to 87)

FAS



PEAK	RT	REL	RT	% CONC	AREA	COMMENT
1	0.405	F	0.22	1.1%	28346	
2	0.560	F	0.30	1.8%	45028	
3	0.888	F	0.48	18.9%	475956	Acetylated F peak
4	1.257	F	0.67	0.7%	17220	
5	1.763	F	0.95	67.9%	1713000	Consistent with F
6	2.608	F	1.40	2.7%	67901	
7	3.627	A	1.00	4.0%	101341	A peak
8	5.127	S	0.99	3.0%	75869	Consistent with S
Total Area: 2524661						A/V=1.33

Codes: 1) Wide A peak
2) Area of A peak < 80%
3) Peak area greater than expected
4) Peak after A2
5) A1c > 10%
6) HbF or variant present
7) Total sample area too small/big
8) A2 is not within normal range