


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 1870, Rack A, Plate 1, Well A02, 456686
[DD7060B0A7A4D39D] Jul 23, 2025 09:20:25 Pressure = 61 bar (61 to 62)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.407	F 0.49	2.0%	33702	
2	0.490	F 0.59	1.4%	23937	
3	0.703	F 0.84	17.0%	281815	Acetylated F peak
4	0.828	F 0.99	66.2%	1100180	Consistent with F
5	1.215	A 1.01	13.3%	221139	A peak
Total Area:				1660773	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range