


LABORATORIO PREGEN  
Carrera 15a No 106-42  
BOGOTA

Batch 2179, Rack A, Plate 1, Well B10, 485902  
[DD6DAA840038E4CA] Jun 18, 2026 13:16:24 Pressure = 70 bar (70 to 72)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.510	F 0.61	2.1%	25738	
2	0.693	F 0.83	16.0%	201274	Acetylated F peak
3	0.838	F 1.00	60.7%	761738	Consistent with F
4	1.188	A 1.00	21.2%	266582	A peak - REVIEW
Total Area: 1255332					

- Codes:
- 1) Wide A peak
  - 2) Area of A peak < 80%
  - 3) Peak area greater than expected
  - 4) Peak after A2
  - 5) Alc > 10%
  - 6) HbF or variant present
  - 7) Total sample area too small/big
  - 8) A2 is not within normal range