


LABORATORIO PREGEN
 Carrera 15a No 106-42
 BOGOTA

Batch 2025, Rack A, Plate 1, Well C09, 171617
 [9863C438BBAAAC68] Dec 23, 2025 12:09:41 Pressure = 76 bar (76 to 78)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.407	F 0.48	1.2%	8977	
2	0.497	F 0.58	1.0%	7215	
3	0.708	F 0.83	13.7%	103486	Acetylated F peak
4	0.845	F 0.99	54.1%	407305	Consistent with F
5	1.002	F 1.18	2.3%	17627	
6	1.225	A 1.00	27.7%	208180	A peak - REVIEW
Total Area:				752790	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range