


LABORATORIO PREGEN
 Carrera 15a No 106-42
 BOGOTA

Batch 1960, Rack A, Plate 1, Well B02, 482280
 [9833F1A5F8D00E40] Oct 28, 2025 10:02:27 Pressure = 67 bar (67 to 70)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.398	F 0.46	1.0%	10712	
2	0.502	F 0.59	0.7%	6902	
3	0.715	F 0.83	17.9%	184538	
4	0.848	F 0.99	58.2%	600825	Consistent with F
5	1.013	F 1.18	2.0%	20241	
6	1.238	A 1.01	20.2%	208699	A peak - REVIEW
Total Area:				1031917	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range