


LABORATORIO PREGEN  
Carrera 15a No 106-42  
BOGOTA

Batch 2140, Rack A, Plate 1, Well H09, 173290  
[8C734B6ACC9C0E48] May 06, 2026 14:39:45 Pressure = 63 bar (62 to 64)

AF



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.410	F 0.48	1.9%	12980	
2	0.498	F 0.58	2.2%	14993	
3	0.742	F 0.86	6.2%	42103	Acetylated F peak
4	0.863	F 1.00	26.1%	178115	Consistent with F
5	1.038	F 1.21	2.0%	13958	
6	1.242	A 1.01	60.4%	412460	A peak - REVIEW
7	1.398	A 1.14	1.3%	8809	3
Total Area:				683418	

- Codes:
- 1) Wide A peak
  - 2) Area of A peak < 80%
  - 3) Peak area greater than expected
  - 4) Peak after A2
  - 5) Alc > 10%
  - 6) HbF or variant present
  - 7) Total sample area too small/big
  - 8) A2 is not within normal range