


LABORATORIO PREGEN  
Carrera 15a No 106-42  
BOGOTA

Batch 2046, Rack A, Plate 1, Well A05, 488969  
[CC328C6C393FF91D] Jan 16, 2026 10:10:09 Pressure = 83 bar (82 to 85)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.407	F 0.49	0.8%	6075	
2	0.503	F 0.60	0.6%	4530	
3	0.698	F 0.83	17.5%	131125	Acetylated F peak
4	0.833	F 0.99	64.2%	479910	Consistent with F
5	0.990	F 1.18	1.8%	13103	
6	1.207	A 1.02	15.1%	112875	A peak
Total Area:				747618	

- Codes:
- 1) Wide A peak
  - 2) Area of A peak < 80%
  - 3) Peak area greater than expected
  - 4) Peak after A2
  - 5) Alc > 10%
  - 6) HbF or variant present
  - 7) Total sample area too small/big
  - 8) A2 is not within normal range