


LABORATORIO PREGEN  
 Carrera 15a No 106-42  
 BOGOTA

Batch 1953, Rack A, Plate 1, Well D02, 859808  
 [98214F9F95698E68] Oct 21, 2025 11:32:05 Pressure = 66 bar (65 to 68)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.397	F 0.46	1.5%	14395	
2	0.508	F 0.59	0.4%	4246	
3	0.635	F 0.74	0.3%	2624	
4	0.742	F 0.87	21.2%	209873	
5	0.845	F 0.99	69.4%	685924	Consistent with F
6	1.002	F 1.17	0.7%	7326	
7	1.235	A 1.00	6.4%	63629	A peak
Total Area:				988017	

- Codes:
- 1) Wide A peak
  - 2) Area of A peak < 80%
  - 3) Peak area greater than expected
  - 4) Peak after A2
  - 5) Alc > 10%
  - 6) HbF or variant present
  - 7) Total sample area too small/big
  - 8) A2 is not within normal range