


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 1870, Rack A, Plate 1, Well A01, 169206
[887135E5F2F10448] Jul 23, 2025 09:18:25 Pressure = 61 bar (61 to 63)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.402	F 0.48	1.3%	8655	
2	0.497	F 0.59	1.2%	8563	
3	0.698	F 0.84	14.1%	96405	Acetylated F peak
4	0.853	F 1.02	64.5%	442218	Consistent with F
5	1.217	A 1.01	18.9%	129532	A peak
Total Area:				685373	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range