


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 1824, Rack A, Plate 1, Well B09, 785461
[D834D1E42502DBBF] Jun 04, 2025 11:21:32 Pressure = 57 bar (57 to 59)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.393	F 0.47	1.3%	12568	
2	0.490	F 0.59	1.0%	9491	
3	0.730	F 0.87	16.1%	159489	Acetylated F peak
4	0.838	F 1.00	64.5%	638692	Consistent with F
5	1.225	A 1.02	17.1%	169592	A peak
Total Area:				989832	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range