


LABORATORIO PREGEN
 Carrera 15a No 106-42
 BOGOTA

Batch 2113, Rack A, Plate 1, Well H04, 498289
 [8C2606932823DB17] Apr 07, 2026 14:09:54 Pressure = 59 bar (58 to 60)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.423	F 0.47	1.1%	11764	
2	0.512	F 0.57	0.8%	8909	
3	0.730	F 0.82	12.4%	133233	Acetylated F peak
4	0.885	F 0.99	60.9%	655065	Consistent with F
5	1.102	A 0.88	0.2%	2634	
6	1.277	A 1.02	24.5%	263155	A peak - REVIEW
Total Area:				1074760	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range