


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 2121, Rack A, Plate 1, Well E10, 502322
[992646565621FB9F] Apr 13, 2026 12:28:40 Pressure = 57 bar (57 to 59)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.512	F 0.59	3.0%	42338	
2	0.725	F 0.83	14.4%	206788	Acetylated F peak
3	0.877	F 1.00	66.4%	951667	Consistent with F
4	1.240	A 1.00	16.3%	233364	A peak
Total Area:				1434157	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range