


LABORATORIO PREGEN  
Carrera 15a No 106-42  
BOGOTA

Batch 2173, Rack A, Plate 1, Well A02, 173605  
[9C69BD278667B13D] Jun 10, 2026 13:03:33 Pressure = 68 bar (67 to 69)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.503	F 0.60	5.4%	32920	
2	0.695	F 0.83	16.2%	99115	Acetylated F peak
3	0.838	F 1.00	58.8%	360433	Consistent with F
4	1.197	A 1.01	19.6%	120054	A peak

Total Area: 612522

- Codes:
- 1) Wide A peak
  - 2) Area of A peak < 80%
  - 3) Peak area greater than expected
  - 4) Peak after A2
  - 5) Alc > 10%
  - 6) HbF or variant present
  - 7) Total sample area too small/big
  - 8) A2 is not within normal range