


LABORATORIO PREGEN
 Carrera 15a No 106-42
 BOGOTA

Batch 1905, Rack A, Plate 1, Well D02, 170105
 [887041B78509AE4A] Sep 10, 2025 12:13:36 Pressure = 61 bar (61 to 63)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.405	F 0.48	0.3%	7809	
2	0.503	F 0.59	0.3%	7363	
3	0.695	F 0.82	13.8%	327594	Acetylated F peak
4	0.820	F 0.96	69.8%	1652245	Consistent with F
5	1.198	A 1.01	15.7%	372571	A peak
Total Area:				2367582	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range