


LABORATORIO PREGEN  
Carrera 15a No 106-42  
BOGOTA

Batch 2113, Rack A, Plate 1, Well H11, 502306  
[9C7757C67DD62EEA] Apr 07, 2026 14:26:04 Pressure = 58 bar (58 to 60)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.423	F 0.47	1.0%	13967	
2	0.512	F 0.56	0.7%	9880	
3	0.728	F 0.80	10.9%	156434	Acetylated F peak
4	0.870	F 0.96	62.0%	892183	Consistent with F
5	1.265	A 1.01	25.5%	367547	A peak - REVIEW

Total Area: 1440011

- Codes:
- 1) Wide A peak
  - 2) Area of A peak < 80%
  - 3) Peak area greater than expected
  - 4) Peak after A2
  - 5) Alc > 10%
  - 6) HbF or variant present
  - 7) Total sample area too small/big
  - 8) A2 is not within normal range