


LABORATORIO PREGEN  
Carrera 15a No 106-42  
BOGOTA

Batch 2188, Rack A, Plate 1, Well A04, 486245  
[896891D300DD1395] Jul 01, 2026 10:58:35 Pressure = 79 bar (79 to 81)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.527	F 0.63	0.8%	16093	
2	0.685	F 0.82	15.7%	309050	Acetylated F peak
3	0.813	F 0.97	75.2%	1482155	Consistent with F
4	1.167	A 1.00	8.3%	164494	A peak

Total Area: 1971792

- Codes:
- 1) Wide A peak
  - 2) Area of A peak < 80%
  - 3) Peak area greater than expected
  - 4) Peak after A2
  - 5) Alc > 10%
  - 6) HbF or variant present
  - 7) Total sample area too small/big
  - 8) A2 is not within normal range