


LABORATORIO PREGEN  
Carrera 15a No 106-42  
BOGOTA

Batch 1911, Rack A, Plate 1, Well D02, 170204  
[8C71006DE4D4D91D] Sep 17, 2025 13:18:57 Pressure = 60 bar (59 to 61)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.403	F 0.47	1.0%	12297	
2	0.497	F 0.58	0.6%	7183	
3	0.698	F 0.82	12.4%	151438	Acetylated F peak
4	0.828	F 0.97	68.8%	842064	Consistent with F
5	1.203	A 1.01	17.3%	211673	A peak
Total Area:				1224655	

- Codes:
- 1) Wide A peak
  - 2) Area of A peak < 80%
  - 3) Peak area greater than expected
  - 4) Peak after A2
  - 5) Alc > 10%
  - 6) HbF or variant present
  - 7) Total sample area too small/big
  - 8) A2 is not within normal range