


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 2173, Rack A, Plate 1, Well A08, 502240
[DC29F92786CDBBBF] Jun 10, 2026 13:15:40 Pressure = 67 bar (67 to 69)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.518	F 0.62	1.3%	12334	
2	0.687	F 0.82	16.3%	153902	Acetylated F peak
3	0.833	F 0.99	69.4%	655033	Consistent with F
4	1.198	A 1.01	13.0%	123048	A peak

Total Area: 944317

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range