


LABORATORIO PREGEN  
Carrera 15a No 106-42  
BOGOTA

Batch 2046, Rack A, Plate 1, Well A10, 488975  
[982399396C6006E0] Jan 16, 2026 10:20:15 Pressure = 83 bar (83 to 85)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.412	F 0.49	0.8%	8508	
2	0.500	F 0.60	0.7%	7217	
3	0.702	F 0.84	15.6%	160408	Acetylated F peak
4	0.842	F 1.00	63.8%	657000	Consistent with F
5	1.200	A 1.01	19.1%	196785	A peak
Total Area:				1029918	

- Codes:
- 1) Wide A peak
  - 2) Area of A peak < 80%
  - 3) Peak area greater than expected
  - 4) Peak after A2
  - 5) Alc > 10%
  - 6) HbF or variant present
  - 7) Total sample area too small/big
  - 8) A2 is not within normal range