


LABORATORIO PREGEN  
Carrera 15a No 106-42  
BOGOTA

Batch 1897, Rack A, Plate 1, Well A03, 787792  
[C9642EF178312C6A] Aug 28, 2025 08:19:16 Pressure = 64 bar (63 to 66)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.402	F 0.48	2.1%	13067	
2	0.488	F 0.58	2.8%	17345	
3	0.673	F 0.80	12.4%	75793	Acetylated F peak
4	0.837	F 1.00	51.1%	313123	Consistent with F
5	0.982	F 1.17	3.7%	22467	
6	1.212	A 1.01	27.3%	167475	A peak - REVIEW
7	1.357	A 1.13	0.6%	3728	
Total Area:				612998	

- Codes:
- 1) Wide A peak
  - 2) Area of A peak < 80%
  - 3) Peak area greater than expected
  - 4) Peak after A2
  - 5) Alc > 10%
  - 6) HbF or variant present
  - 7) Total sample area too small/big
  - 8) A2 is not within normal range