


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 2182, Rack A, Plate 1, Well F07, 485941
[8C2DBA68BB4DBB9D] Jun 23, 2026 12:26:33 Pressure = 71 bar (71 to 73)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.510	F 0.61	2.4%	23007	
2	0.690	F 0.82	14.5%	141170	Acetylated F peak
3	0.832	F 0.99	53.5%	520456	Consistent with F
4	0.982	F 1.17	2.7%	26406	
5	1.195	A 1.02	26.6%	258489	A peak - REVIEW
6	1.367	S 0.88	0.3%	3334	
Total Area:				972862	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range