


LABORATORIO PREGEN  
 Carrera 15a No 106-42  
 BOGOTA

Batch 2107, Rack A, Plate 1, Well E03, 498862  
 [8C6214D4025C2660] Mar 25, 2026 12:54:47 Pressure = 57 bar (56 to 58)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.508	F 0.58	6.2%	45429	
2	0.728	F 0.83	5.5%	40538	
3	0.787	F 0.90	1.3%	9826	
4	0.868	F 0.99	74.5%	545464	Consistent with F
5	1.262	A 1.01	12.4%	91110	A peak
Total Area:				732367	

- Codes:
- 1) Wide A peak
  - 2) Area of A peak < 80%
  - 3) Peak area greater than expected
  - 4) Peak after A2
  - 5) Alc > 10%
  - 6) HbF or variant present
  - 7) Total sample area too small/big
  - 8) A2 is not within normal range