


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 1953, Rack A, Plate 1, Well B08, 859884
[CC751BCAC2BEDB3D] Oct 21, 2025 10:53:41 Pressure = 66 bar (66 to 68)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.393	F 0.46	1.0%	10555	
2	0.515	F 0.60	0.4%	4582	
3	0.717	F 0.84	10.7%	112352	Acetylated F peak
4	0.847	F 0.99	76.1%	799004	Consistent with F
5	1.238	A 1.01	11.7%	122967	A peak
Total Area:				1049460	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range