


LABORATORIO PREGEN  
 Carrera 15a No 106-42  
 BOGOTA

Batch 1999, Rack A, Plate 1, Well D06, 171239  
 [9D33FD36F110AE60] Nov 28, 2025 14:11:03

Pressure = 72 bar (71 to 75) AF



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.395	F 0.20	3.8%	59950	
2	0.525	F 0.27	3.0%	47450	
3	0.870	F 0.44	3.7%	57838	
4	1.215	F 0.62	0.3%	4831	
5	1.317	F 0.67	0.2%	3540	
6	1.670	F 0.85	0.7%	10429	Alc or possibly HbH
7	1.975	F 1.00	6.4%	101139	6
8	2.257	F 1.15	0.3%	5249	
9	2.723	F 1.38	3.8%	59622	
10	2.997	A 0.79	2.8%	45044	
11	3.220	A 0.85	0.3%	5370	
12	3.755	A 0.99	72.4%	1145850	2 A peak
13	4.563	S 0.86	2.3%	36008	
Total Area:				1582320	

- Codes:
- 1) Wide A peak
  - 2) Area of A peak < 80%
  - 3) Peak area greater than expected
  - 4) Peak after A2
  - 5) Alc > 10%
  - 6) HbF or variant present
  - 7) Total sample area too small/big
  - 8) A2 is not within normal range