


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 1801, Rack A, Plate 1, Well B03, 455886
[DC65AD12F6E50EC8] May 21, 2025 10:19:37 Pressure = 60 bar (59 to 61)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.407	F 0.48	1.0%	8994	
2	0.498	F 0.59	1.1%	10583	
3	0.698	F 0.83	16.6%	156281	Acetylated F peak
4	0.840	F 0.99	66.5%	627828	Consistent with F
5	1.228	A 1.01	14.8%	140042	A peak
Total Area:				943728	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range