


LABORATORIO PREGEN
 Carrera 15a No 106-42
 BOGOTA

Batch 1911, Rack A, Plate 1, Well C06, 455733
 [9D61046DE474FB35] Sep 17, 2025 13:02:47 Pressure = 61 bar (60 to 61)

FA



| PEAK | RT | REL RT | % CONC | AREA | COMMENT |
|-------------|-------|--------|--------|--------|-------------------|
| 1 | 0.410 | F 0.48 | 1.0% | 10122 | |
| 2 | 0.497 | F 0.58 | 1.1% | 10836 | |
| 3 | 0.683 | F 0.80 | 16.0% | 156490 | Acetylated F peak |
| 4 | 0.828 | F 0.97 | 62.0% | 606191 | Consistent with F |
| 5 | 0.982 | F 1.15 | 1.6% | 15224 | |
| 6 | 1.205 | A 1.01 | 18.1% | 177228 | A peak |
| 7 | 1.357 | A 1.14 | 0.2% | 2021 | |
| Total Area: | | | | 978112 | |

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range