


LABORATORIO PREGEN
 Carrera 15a No 106-42
 BOGOTA

Batch 2025, Rack A, Plate 1, Well C08, 171616
 [8C62916DEEFF79BF] Dec 23, 2025 12:07:40 Pressure = 77 bar (76 to 78)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.407	F 0.48	1.1%	8999	
2	0.497	F 0.58	1.1%	9007	
3	0.708	F 0.83	15.6%	128939	Acetylated F peak
4	0.852	F 1.00	60.3%	499582	Consistent with F
5	0.997	F 1.17	1.2%	9984	
6	1.223	A 1.00	20.8%	172002	A peak - REVIEW
Total Area:				828513	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range