


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 1998, Rack A, Plate 1, Well A08, 171208
[8C33BD67AE65D9BD] Nov 28, 2025 11:02:37 Pressure = 72 bar (71 to 73)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.395	F 0.46	1.6%	20611	
2	0.497	F 0.58	0.5%	7099	
3	0.708	F 0.83	14.5%	186959	Acetylated F peak
4	0.850	F 0.99	66.4%	857607	Consistent with F
5	1.237	A 1.01	16.9%	218467	A peak
Total Area:				1290743	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range