


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 2029, Rack A, Plate 1, Well C09, 490083
[8C3397A890002EE8] Dec 29, 2025 11:58:53 Pressure = 74 bar (74 to 77)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.417	F 0.49	1.2%	14342	
2	0.495	F 0.58	0.7%	8121	
3	0.752	F 0.88	14.4%	174078	Acetylated F peak
4	0.843	F 0.99	66.2%	800083	Consistent with F
5	1.023	F 1.20	1.5%	18694	
6	1.230	A 1.02	16.0%	194008	A peak

Total Area: 1209326

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range