


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 2068, Rack A, Plate 1, Well D07, 492217
[9D77642E6659F99F] Feb 10, 2026 11:50:32 Pressure = 95 bar (95 to 98)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.412	F 0.49	0.7%	9921	
2	0.507	F 0.60	0.4%	5890	
3	0.718	F 0.85	16.5%	239546	Acetylated F peak
4	0.825	F 0.98	70.6%	1022152	Consistent with F
5	0.982	F 1.16	0.6%	8990	
6	1.195	A 1.01	11.2%	162105	A peak
Total Area:				1448604	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range