


LABORATORIO PREGEN  
Carrera 15a No 106-42  
BOGOTA

Batch 2191, Rack A, Plate 1, Well B01, 493039  
[993D952AE65F13BD] Jul 06, 2026 12:04:45 Pressure = 79 bar (79 to 81)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.510	F 0.61	1.2%	10614	
2	0.683	F 0.82	11.6%	103292	Acetylated F peak
3	0.825	F 0.99	62.5%	556865	Consistent with F
4	1.168	A 1.00	24.7%	220248	A peak - REVIEW

Total Area: 891019

- Codes:
- 1) Wide A peak
  - 2) Area of A peak < 80%
  - 3) Peak area greater than expected
  - 4) Peak after A2
  - 5) Alc > 10%
  - 6) HbF or variant present
  - 7) Total sample area too small/big
  - 8) A2 is not within normal range