


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 1953, Rack A, Plate 1, Well D04, 859811
[D9305ACAC036719F] Oct 21, 2025 11:36:08 Pressure = 66 bar (65 to 67)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.397	F 0.46	1.3%	15132	
2	0.500	F 0.58	0.5%	6367	
3	0.738	F 0.86	14.7%	170945	Acetylated F peak
4	0.857	F 1.00	56.8%	658471	Consistent with F
5	0.993	F 1.16	1.0%	11363	
6	1.225	A 0.99	25.7%	297608	A peak - REVIEW

Total Area: 1159886

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range