


LABORATORIO PREGEN
 Carrera 15a No 106-42
 BOGOTA

Batch 2065, Rack A, Plate 1, Well D07, 498927
 [D93725EACF862C68] Feb 04, 2026 13:05:09 Pressure = 86 bar (85 to 88)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.408	F 0.49	0.6%	8662	
2	0.508	F 0.60	0.4%	5077	
3	0.697	F 0.83	16.3%	222989	Acetylated F peak
4	0.842	F 1.00	71.4%	974313	Consistent with F
5	1.190	A 1.01	11.2%	152973	A peak
Total Area:				1364014	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range