


LABORATORIO PREGEN  
Carrera 15a No 106-42  
BOGOTA

Batch 2110, Rack A, Plate 1, Well C01, 172824  
[DD73557AE4D42E60] Mar 31, 2026 11:50:15 Pressure = 59 bar (59 to 61)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.512	F 0.59	6.4%	32522	
2	0.727	F 0.84	17.6%	89197	
3	0.870	F 1.00	43.2%	219140	Consistent with F
4	1.000	F 1.15	8.1%	41026	3
5	1.248	A 1.01	24.6%	124843	A peak - REVIEW
Total Area:				506728	

- Codes:
- 1) Wide A peak
  - 2) Area of A peak < 80%
  - 3) Peak area greater than expected
  - 4) Peak after A2
  - 5) Alc > 10%
  - 6) HbF or variant present
  - 7) Total sample area too small/big
  - 8) A2 is not within normal range