


- - - - -  
`LABORATORIO PREGEN  
Carrera 15a No 106-42  
BOGOTA

Batch 2204, Rack A, Plate 1, Well B01, 500818  
[C8698DF9453F9197] Jul 23, 2026 10:35:22 Pressure = 92 bar (92 to 94)

FA



| PEAK                | RT    | REL | RT   | % CONC | AREA   | COMMENT           |
|---------------------|-------|-----|------|--------|--------|-------------------|
| 1                   | 0.692 | F   | 0.84 | 23.3%  | 332394 |                   |
| 2                   | 0.823 | F   | 0.99 | 62.2%  | 888367 | Consistent with F |
| 3                   | 1.158 | A   | 1.00 | 14.5%  | 207622 | A peak            |
| Total Area: 1428383 |       |     |      |        |        |                   |

Codes: 1) Wide A peak  
2) Area of A peak < 80%  
3) Peak area greater than expected  
4) Peak after A2  
5) A1c > 10%  
6) HbF or variant present  
7) Total sample area too small/big  
8) A2 is not within normal range