


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 2130, Rack A, Plate 1, Well E03, 502916
[8D63192AC6E37BBD] Apr 24, 2026 12:54:21 Pressure = 61 bar (61 to 63)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.510	F 0.59	3.2%	39735	
2	0.718	F 0.83	15.8%	194138	Acetylated F peak
3	0.858	F 0.99	70.0%	858485	Consistent with F
4	1.225	A 1.00	10.9%	134208	A peak
Total Area:				1226566	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range