


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 2034, Rack A, Plate 1, Well D12, 490359
[CD7397467E02AE42] Jan 06, 2026 12:13:38 Pressure = 81 bar (81 to 83)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.418	F 0.49	0.6%	10981	
2	0.503	F 0.59	0.4%	6661	
3	0.702	F 0.82	16.0%	288568	Acetylated F peak
4	0.830	F 0.97	71.2%	1285682	Consistent with F
5	1.200	A 1.01	11.9%	214721	A peak
Total Area:				1806613	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range