


LABORATORIO PREGEN
 Carrera 15a No 106-42
 BOGOTA

Batch 2124, Rack A, Plate 1, Well G08, 502343
 [D8731CF652437B3D] Apr 16, 2026 12:54:13 Pressure = 62 bar (61 to 63)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.432	F 0.49	1.4%	13372	
2	0.512	F 0.59	0.9%	9052	
3	0.713	F 0.82	16.3%	161787	Acetylated F peak
4	0.852	F 0.98	63.0%	623608	Consistent with F
5	0.998	F 1.14	1.7%	16958	
6	1.242	A 0.99	16.7%	165638	A peak
Total Area:				990415	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range