


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 2061, Rack A, Plate 1, Well F06, 172059
[88768E05F635DB95] Jan 30, 2026 12:16:59

Pressure = 88 bar (88 to 90)

FA



PEAK	RT	REL	RT	% CONC	AREA	COMMENT
1	0.493	F	0.58	2.1%	20363	
2	0.753	F	0.89	13.5%	129153	Acetylated F peak
3	0.838	F	0.99	59.1%	566566	Consistent with F
4	1.010	F	1.20	1.3%	12589	
5	1.207	A	1.01	23.8%	227817	A peak - REVIEW
6	1.387	S	0.89	0.2%	2320	
Total Area: 958808						

Codes: 1) Wide A peak
2) Area of A peak < 80%
3) Peak area greater than expected
4) Peak after A2
5) A1c > 10%
6) HbF or variant present
7) Total sample area too small/big
8) A2 is not within normal range