


- - - - -
`LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 2204, Rack A, Plate 1, Well G07, 500831
[D8698CBC386AE44A] Jul 23, 2026 12:52:48 Pressure = 93 bar (93 to 96)

FA



PEAK	RT	REL	RT	% CONC	AREA	COMMENT
1	0.683	F	0.81	14.5%	183765	Acetylated F peak
2	0.820	F	0.97	72.0%	909573	Consistent with F
3	1.150	A	1.00	13.5%	170420	A peak
Total Area: 1263758						

Codes: 1) Wide A peak
2) Area of A peak < 80%
3) Peak area greater than expected
4) Peak after A2
5) A1c > 10%
6) HbF or variant present
7) Total sample area too small/big
8) A2 is not within normal range