


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 2143, Rack A, Plate 1, Well A05, 496509
[DC775B1C1EB4A4E8] May 08, 2026 11:31:57 Pressure = 64 bar (64 to 66)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.513	F 0.60	1.5%	12947	
2	0.712	F 0.84	14.0%	120426	Acetylated F peak
3	0.858	F 1.01	65.8%	567382	Consistent with F
4	1.217	A 1.01	18.8%	162125	A peak

Total Area: 862880

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range