


LABORATORIO PREGEN  
 Carrera 15a No 106-42  
 BOGOTA

Batch 1988, Rack A, Plate 1, Well D08, 482021  
 [9D66A9B5D8EF513D] Nov 21, 2025 12:07:01 Pressure = 68 bar (68 to 70)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.398	F 0.46	0.9%	11071	
2	0.497	F 0.58	0.8%	10273	
3	0.730	F 0.85	16.2%	206954	Acetylated F peak
4	0.847	F 0.99	68.9%	878223	Consistent with F
5	1.017	F 1.18	1.0%	13096	
6	1.240	A 1.01	12.2%	154894	A peak
Total Area:				1274511	

- Codes:
- 1) Wide A peak
  - 2) Area of A peak < 80%
  - 3) Peak area greater than expected
  - 4) Peak after A2
  - 5) Alc > 10%
  - 6) HbF or variant present
  - 7) Total sample area too small/big
  - 8) A2 is not within normal range