


LABORATORIO PREGEN
 Carrera 15a No 106-42
 BOGOTA

Batch 1911, Rack A, Plate 1, Well C02, 455711
 [9D751138B12926CA] Sep 17, 2025 12:52:40 Pressure = 60 bar (60 to 61)

FA



| PEAK | RT | REL RT | % CONC | AREA | COMMENT |
|-------------|-------|--------|--------|--------|-------------------|
| 1 | 0.408 | F 0.49 | 1.4% | 12144 | |
| 2 | 0.493 | F 0.59 | 1.3% | 10795 | |
| 3 | 0.688 | F 0.83 | 10.9% | 93537 | Acetylated F peak |
| 4 | 0.832 | F 1.00 | 52.4% | 450598 | Consistent with F |
| 5 | 0.977 | F 1.17 | 4.5% | 38673 | |
| 6 | 1.203 | A 1.01 | 28.7% | 246894 | A peak - REVIEW |
| 7 | 1.362 | A 1.14 | 0.9% | 7649 | |
| Total Area: | | | | 860290 | |

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range