


- - - - -
`LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 2191, Rack A, Plate 1, Well G01, 493203
[9D2DC42AE4FD9B95] Jul 06, 2026 14:10:03 Pressure = 78 bar (78 to 80)

AF



PEAK	RT	REL	RT	% CONC	AREA	COMMENT
1	0.415	F	0.50	1.7%	8158	
2	0.500	F	0.60	1.0%	5037	
3	0.683	F	0.82	4.8%	23444	Acetylated F peak
4	0.843	F	1.01	37.3%	180682	Consistent with F
5	1.175	A	1.01	55.1%	266711	A peak - REVIEW
Total Area:					484032	

Codes: 1) Wide A peak
2) Area of A peak < 80%
3) Peak area greater than expected
4) Peak after A2
5) A1c > 10%
6) HbF or variant present
7) Total sample area too small/big
8) A2 is not within normal range