


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 2124, Rack A, Plate 1, Well C11, 502330
[8C630DB65AEB7197] Apr 16, 2026 11:21:14 Pressure = 62 bar (62 to 63)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.510	F 0.58	2.8%	35451	
2	0.733	F 0.83	16.9%	217054	Acetylated F peak
3	0.850	F 0.97	66.0%	849437	Consistent with F
4	1.243	A 1.01	14.4%	184783	A peak
Total Area: 1286725					

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range